



PicoMaxx High Fidelity PCR System

Instruction Manual

Catalog #600420 (100 U), #600422 (500 U), and #600424 (1000 U)

Revision B

Research Use Only. Not for Use in Diagnostic Procedures.

600420-12



Agilent Technologies

LIMITED PRODUCT WARRANTY

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

ORDERING INFORMATION AND TECHNICAL SERVICES

United States and Canada

Agilent Technologies
Stratagene Products Division

11011 North Torrey Pines Road
La Jolla, CA 92037

Telephone (858) 373-6300
Order Toll Free (800) 424-5444
Technical Services (800) 894-1304
Email techservices@agilent.com
World Wide Web www.genomics.agilent.com

Europe

Location	Telephone
Austria	01 25125 6800
Benelux	02 404 92 22
Denmark	45 70 13 00 30
Finland	010 802 220
France	0810 446 446
Germany	0800 603 1000
Italy	800 012575
Netherlands	020 547 2600
Spain	901 11 68 90
Sweden	08 506 4 8960
Switzerland	0848 8035 60
UK/Ireland	0845 712 5292

All Other Countries

Please contact your local distributor. A complete list of distributors is available at www.genomics.agilent.com.

PicoMaxx High Fidelity PCR System

CONTENTS

Materials Provided	1
Storage Conditions	1
Additional Materials Required	1
Notices to Purchaser	2
Introduction	3
Optimization Parameters	3
DNA Template Quality and Concentration	3
Primer Design and Concentration	4
Deoxynucleotide Concentrations.....	4
Enzyme Concentration	4
Reaction Buffer	4
Cycling Parameters.....	4
PCR Protocol	5
Troubleshooting	7
References	7
MSDS Information	7
Quick-Reference Protocol	10

PicoMaxx High Fidelity PCR System

MATERIALS PROVIDED

Materials provided	Quantity		
	Catalog #600420	Catalog #600422	Catalog #600424
PicoMaxx High Fidelity PCR System (2.5 U/ μ l)	100 U	500 U	1000 U
10× PicoMaxx Reaction Buffer ^a	1 ml	2 × 1 ml	4 × 1 ml

^a The total Mg²⁺ concentration present in the final 1× dilution of the 10× PicoMaxx reaction buffer is 2 mM.

STORAGE CONDITIONS

All Components: -20°C

ADDITIONAL MATERIALS REQUIRED

Temperature cycler
PCR tubes^{ll}
PCR primers
Deoxynucleotides

^{ll} Thin-walled PCR tubes are highly recommended for use with Stratagene RoboCycler thermal cyclers. These PCR tubes are optimized to ensure ideal contact with the multiblock design to permit more efficient heat transfer and to maximize thermal-cycling performance.

NOTICES TO PURCHASER

Limited Label License for *Pfu*-Containing DNA Polymerase Products

This product is covered by the claims of one or more of the following U.S. Patents: 5,545,552; 5,556,772; 5,866,395; 5,948,663; 6,489,150; 6,183,997; 6,444,428, and patents pending. Purchase of this product conveys to the purchaser only the non-transferable right under these patents to use the product ***for research use only by the purchaser***. No rights are granted to the purchaser hereunder to sell, modify for resale or otherwise transfer this product, either alone or as a component of another product, to any third party. Agilent reserves all other rights, and this product may not be used in any manner other than as provided herein. For information on obtaining a license to use this product for purposes other than research, please contact the Stratagene Products Division, Business Development, 11011 North Torrey Pines Road, La Jolla, California 92037. Phone (858) 535-5400.

Notice to Purchaser: Limited License

Purchase of this product includes an immunity from suit under patents specified in the product insert to use only the amount purchased for the purchaser's own internal research. No other patent rights (such as 5' Nuclease Process patent rights) are conveyed expressly, by implication, or by estoppel. Further information on purchasing licenses may be obtained by contacting the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

INTRODUCTION

The PicoMaxx high fidelity PCR system* is a blend of cloned *Taq* and *Pfu* DNA polymerases,** and Stratagene ArchaeMaxx polymerase-enhancing factor. Together with a specially optimized buffer, this enzyme blend provides the highest success rate of any PCR enzyme, even higher than other *Taq*-based blends. The PicoMaxx high fidelity PCR system reliably produces high PCR product yields on a wide variety of templates up to 10 kb. It also provides superior sensitivity by successfully amplifying samples where starting material is limited.

The PicoMaxx high fidelity PCR system is formulated with antibodies that inhibit polymerase activity until cycling begins, promoting high-specificity and reduced background. This formulation is ideal for high-throughput, multi-target amplification experiments without requiring optimization solutions or protein enhancers that can decrease PCR efficiency. The PicoMaxx high fidelity PCR system detects low-copy-number targets in complex DNA backgrounds, amplifies difficult systems with minimal non-specific amplification, and works with PCR applications that require prolonged incubations at room temperature prior to thermal cycling.

OPTIMIZATION PARAMETERS

The PicoMaxx high fidelity PCR system reliably produces high PCR product yields on a wide variety of templates up to 10 kb. The PicoMaxx high fidelity PCR system can be introduced into all existing amplification systems optimized with *Taq* or *Taq2000* DNA polymerase, with only slight optimization of reaction and cycling conditions. For other systems, optimization may be required to achieve highest product yield and specificity. Critical optimization parameters include the quality and concentration of DNA template, primers, and deoxynucleotides; primer design; the concentration of enzyme; use of the recommended reaction buffer; and cycling parameters.

DNA Template Quality and Concentration

Successful amplification is dependent upon the purity, integrity, concentration, and molecular weight of the DNA template. Isolation of intact, high molecular weight genomic DNA may be achieved by using the Stratagene DNA Extraction Kit or the RecoverEase DNA isolation kit. Potential shearing of the genomic DNA template is minimized by the use of wide-bore tips for pipetting or mixing of the template. Additionally, freezing of high molecular weight templates should be avoided; storage at 4°C is recommended. The length of an intact genomic DNA template should be >50 kb.

* U.S. Patent Nos. 6,734,293, 6,489,150, 6,444,428, 6,183,997, 5,948,663, 5,866,395, 5,556,772, 5,545,552 and patents pending.

** U.S. Patent Nos. 7,045,328, 6,489,150, 5,948,663, 5,866,395 and 5,545,552, and patents pending.

For amplifying genomic DNA templates, use 100–200 ng of template. To amplify low-complexity targets (for example, lambda DNA or cloned DNA), use 5–20 ng of template. Excess template DNA can inhibit the PCR reaction.

Primer Design and Concentration

Primers should be ≥ 20 bp in length with a balanced $T_m \geq 60^\circ\text{C}$. The resulting high annealing temperature promotes specificity and discourages secondary structure formation. Further, primer sequences should be analyzed for potential duplex and hairpin formation as well as false priming sites in order to obtain the highest yield of specific PCR products.

We suggest using primers at a final concentration of 0.3–0.5 μM , which is equivalent to ~ 100 –250 ng of a 20- to 30-mer oligonucleotide primer in a 50- μl reaction volume.

Deoxynucleotide Concentrations

Amplification efficiencies are influenced by deoxynucleotide (dNTP) concentrations. Insufficient concentrations of dNTPs may result in lower yields. Generally, 200 μM of each dNTP is optimal. For targets ≥ 5 kb, the dNTP concentration may need to be increased to 500 μM each.

Enzyme Concentration

Robust product yield requires an adequate DNA polymerase concentration. The use of 2.5 U/50- μl reaction consistently generates high yields of PCR product. Longer templates may require 5 U of PicoMaxx enzyme per 50- μl reaction for optimal results.

Reaction Buffer

For optimal yield and specificity it is essential that the provided PCR buffer is used. Suboptimal results will be achieved using other buffers. The provided 10 \times PicoMaxx reaction buffer contains the magnesium ion concentration that is optimal for the enzyme. Adjusting the magnesium concentration is not recommended.

Cycling Parameters

As with all PCR reactions, cycling parameters are critical for successful amplification and may require further optimization.

Extension Time

For general applications, maintain an extension time of 1.0 minute/kb of template. Longer extension times of up to 90 seconds/kb of PCR target may produce higher yields for difficult targets, such as high-complexity targets or targets > 2 kb in length.

Annealing Temperature

If using an existing primer pair, the annealing temperature may require optimization in the 10 \times PicoMaxx reaction buffer provided.

PCR PROTOCOL

1. Prepare a reaction mixture for the appropriate number of samples to be amplified. Add the following components *in order* and mix gently. Table I provides an example of a reaction mixture for the amplification of a typical single-copy chromosomal target. The recipe listed in Table I is for one reaction and must be adjusted for multiple samples. The final volume of each sample reaction is 50 µl.

Note *If the tube of 10× PicoMaxx reaction buffer contains a visible precipitate, heat the tube to 37°C for 2–3 minutes to redissolve the contents. Vortex briefly to mix. This heating step does not impact PCR results.*

TABLE I

Reaction Mixture for a Typical Single-Copy Chromosomal Locus PCR Amplification

Component	Amount per reaction
Distilled water (dH ₂ O)	40.6 µl
10× PicoMaxx reaction buffer ^a	5.0 µl
100 mM dNTP mix (25 mM of each dNTP) ^b	0.4 µl
DNA template (100 ng/µl) ^c	1.0 µl
Primer #1 (100 ng/µl) ^{d,e}	1.0 µl
Primer #2 (100 ng/µl) ^{d,e}	1.0 µl
PicoMaxx high fidelity PCR system	1.0 µl (2.5 U)
Total reaction volume	50 µl

^a The total Mg²⁺ concentration present in the final 1× dilution of the 10× PicoMaxx reaction buffer is 2 mM.

^b For targets ≥5 kb, the dNTP concentration may need to be increased to 500 µM each.

^c The amount of DNA template required will vary depending on the type of DNA being amplified. Generally 100–200 ng of genomic DNA template is recommended. Less DNA template (typically 5–20 ng) can be used for amplification of lambda or plasmid PCR targets.

^d Primer concentrations between 0.3 and 0.5 µM are recommended (generally 100–250 ng for typical 20- to 30-mer oligonucleotide primers in a 50-µl reaction volume).

^e For best results, design primers to have similar melting temperatures (typically between 60° and 80°C).

2. Immediately before thermal cycling, aliquot 50 µl of the reaction mixture into the appropriate number of sterile 0.2-ml or 0.5-ml thin-wall PCR tubes or standard microcentrifuge tubes.
3. If the temperature cycler is not equipped with a heated cover, overlay each reaction with ~50 µl of DNase-, RNase-, and protease-free mineral oil (Sigma, St. Louis, Missouri).
4. Perform PCR using the optimized cycling conditions in Table II.

TABLE II
PCR Cycling Parameters^{a,b}

Segment	Number of cycles	Temperature	Duration
1	1	95°C	2 minutes
2	30–35	95°C	40 seconds
		Primer $T_m - 5^\circ\text{C}^c$	30 seconds
		72°C	1 minute/kb PCR target
3	1	72°C	10 minutes

^a Thin-wall PCR tubes are highly recommended. These PCR tubes are optimized to ensure more efficient heat transfer and to maximize thermal-cycling performance.

^b Optimized cycling parameters are not necessarily transferable between thermal cyclers designed by different manufacturers; therefore, each manufacturer's recommendations for optimal cycling parameters should be consulted.

^c If using existing primer pairs, the annealing temperatures may require optimization in the 10× PicoMaxx reaction buffer provided.

- Analyze the PCR amplification products on a 1.0–4.0% (w/v) agarose gel by electrophoresis.

TROUBLESHOOTING

Observation	Suggestion(s)
No product or low yield	<p>Ensure that 10× PicoMaxx reaction buffer is used.</p> <p>Increase the amount of PicoMaxx enzyme up to 5 U per 50-μl PCR reaction.</p> <p>Increase extension time to 90 seconds per kb of PCR target.</p> <p>Use intact and highly purified DNA templates.</p> <p>Increase the amount of full-length intact DNA template, adjust the ratio of primer versus template, and/or increase the number of cycles up to a maximum of 40 cycles to optimize yield of the desired product.</p> <p>Remove extraneous salts from the PCR primers and DNA preparations.</p> <p>For targets ≥5 kb, increase the dNTP concentration to 500 μM each.</p> <p>Check the melting temperature, purity, GC content, and length of the primers.</p> <p>Use the recommended primer concentrations between 0.3 and 0.5 μM (generally 100–250 ng for typical 20- to 30-mer oligonucleotide primers in a 50-μl reaction volume).</p> <p>Primer pairs with matching primer melting temperatures (T_m) and complete complementarity between the primer and template are recommended.</p> <p>If using existing primer pairs, the annealing temperatures may require optimization in the 10× PicoMaxx reaction buffer provided.</p> <p>Denaturation times of 30–60 seconds at 94–95°C are usually sufficient, while longer denaturation times may damage the DNA template. Use the shortest denaturation time compatible with successful PCR on the thermal cycler.</p> <p>Use thin-wall PCR tubes. These PCR tubes are optimized to ensure ideal contact with the block to permit more efficient heat transfer and to maximize thermal-cycling performance.</p>
Artifactual smears	<p>Increase the annealing temperature in 4°C increments.</p> <p>Optimize the cycling parameters specifically for the primer-template set and the thermal cycler used.</p>
Multiple bands	<p>Increase the annealing temperature in 4°C increments.</p> <p>Optimize the cycling parameters specifically for the primer-template set and the thermal cycler used.</p> <p>Use Perfect Match PCR enhancer to improve PCR product specificity.</p>

REFERENCES

1. Innis, M. A., Gelfand, D. H., Sninsky, J. J. and White, T. J. (1990). *PCR Protocols: A Guide to Methods and Applications*. Academic Press, New York.

MSDS INFORMATION

The Material Safety Data Sheet (MSDS) information for Stratagene products is provided on the web at <http://www.genomics.agilent.com>. MSDS documents are not included with product shipments.

STRATAGENE

An Agilent Technologies Division

PicoMaxx High Fidelity PCR System

Catalog #600420, 600422, 600424

QUICK-REFERENCE PROTOCOL

- ◆ Prepare reaction mixtures according to the table below, mix gently, and place in thin-walled PCR tubes.

Component	Amount per reaction
Distilled water (dH ₂ O)	40.6 µl
10× PicoMaxx reaction buffer	5.0 µl
100 mM dNTP mix (25 mM of each dNTP)	0.4 µl
DNA template (100 ng/µl)	1.0 µl
Primer #1 (100 ng/µl)	1.0 µl
Primer #2 (100 ng/µl)	1.0 µl
PicoMaxx high fidelity PCR system	1.0 µl (2.5 U)
Total reaction volume	50 µl

- ◆ If the temperature cycler is not equipped with a heated cover, overlay each reaction mixture with 50 µl mineral oil.
- ◆ Perform PCR using the cycling conditions according to the following table:

Segment	Number of cycles	Temperature	Duration
1	1	95°C	2 minutes
2	30–35	95°C	40 seconds
		Primer T _m – 5°C	30 seconds
		72°C	1 minute/kb PCR target
3	1	72°C	10 minutes

- ◆ Analyze the PCR amplification products by gel electrophoresis.